



Date_	20221012	
Time	16.13	h
INSTRUM	Avance_400A	
PROBHD	Z163739_0080	(
PULPROG	zg30	
TD	131072	
SOLVENT	CDC13	
NS	16	
DS	0	
SWH	11904.762	Hz
FIDRES	0.181652	Hz
AQ	5.5050240	sec
RG	101	
DW	42.000	usec
DE	8.79	usec
TE	298.0	K
D1	0.10000000	sec
TD0	1	
SFO1	400.1524711	MHz
NUC1	1H	
P0	3.33	usec
P1	10.00	usec
PLW1	27.00000000	W

```

F2 - Processing parameters
SI                131072
SF              400.1500371 MHz
WDW                EM
SSB                 0
LB              0.10 Hz
GB                 0
PC              1.00

```

