



-27.61
-78.87
-94.74
-94.92
-95.16
-95.37
-95.62
-95.72
-95.87
-96.13
-96.23
-96.46
-96.51
-96.56
-96.60
-96.62
-96.67
-96.72
-96.77
-96.97
-97.15
-97.29
-97.36
-97.36
-100.51
-104.95
-114.15
-118.15
-118.21
-118.95
-121.30
-122.39
-122.53
-124.62
-124.75
-125.57
-126.19
-127.92
-129.54
-130.63
-131.94
-132.24
-139.34
-178.97
-191.35
-195.09
-195.20
-195.56
-195.80
-196.00
-196.05
-196.36
-196.78
-197.22
-197.25
-197.73
-198.11
-198.92
-199.00
-199.13
-199.25
-199.51
-199.68
-200.12
-200.31
-201.12
-201.16

Current Data Parameters
NAME 400_TK_3BA
EXPNO 21
PROCNO 1

F2 - Acquisition Parameters
Date_ 20260714
Time 17.29 h
INSTRUM Avance
PROBHD Z163739_0222 (
PULPROG zgig
TD 131072
SOLVENT CDCl3
NS 16
DS 4
SWH 90909.094 Hz
FIDRES 1.387163 Hz
AQ 0.7208960 sec
RG 101
DW 5.500 usec
DE 6.50 usec
TE 298.0 K
D1 1.00000000 sec
D11 0.03000000 sec
TD0 1
SFO1 376.4701248 MHz
NUC1 19F
P1 12.00 usec
PLW1 29.43499947 W
SFO2 400.1416006 MHz
NUC2 1H
CPDPRG[2] waltz16
PCPD2 90.00 usec
PLW2 21.11599922 W
PLW12 0.16684000 W

F2 - Processing parameters
SI 65536
SF 376.5077756 MHz
WDW EM
SSB 0
LB 0.30 Hz
GB 0
PC 1.00

